



CERTIFICATION by CUSTODIAN of RECORDS
Tucson Police Department Crime Laboratory

184

I hereby certify that these documents are a true and correct copy of the periodic maintenance* for the Intoxilyzer 5000 Serial Number 64-001226 (#8) maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R9-14-404.A.3, for a 0.080AC solution are 0.070AC to 0.090AC.

Date: 2/23/06 Cal 1 0.079 AC Cal 2 0.080 AC

Date: 3/23/06 Cal 1 N/A AC Cal 2 N/A AC

Date: _____ Cal 1 _____ AC Cal 2 _____ AC

Between the dates of N/A and N/A the instrument was operating properly and accurately.


The instrument was taken out of service for repair on 3/23/06.
The instrument was returned to service on N/A.

Jennifer L. Coombs
Jennifer L. Coombs

3/31/06

ADPS QAS Permit # QAS 29886

SUBSCRIBED AND SWORN to before me this 31, day of March, 2006.



Kathy A. Kinney
Notary Public

*Includes documents specified in A.A.C. R9-14-404.A.6

Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.



TUCSON POLICE DEPARTMENT
CRIME LABORATORY

INTOXILYZER SERVICE RECORD

TS4

Date: March 23, 2006

QAS: Jennifer Coombs 461160

Intoxilyzer Serial Number: 64-001226 (#8)

Reason For Service: Processor Error 1 during initial
diagnostic checks.

Date Out of Service: 3/23/06

Calibration Check completed prior to removal from service

Date: N/A

SQAP completed prior to removal from service

Date: N/A

Corrective Action:

☒ Sent to Applied Electronics



MEMORANDUM

DATE: March 30, 2006

TO: Whom it may concern

FROM: Jennifer L. Coombs
Intoxilyzer 5000 QAS
Tucson Police Department
Crime Lab
270 S. Stone Avenue
Tucson, AZ 85701-1917
(520) 791-4494

A handwritten signature, likely of Jennifer L. Coombs, written in cursive.

SUBJECT: Intoxilyzer Out of Service

Intoxilyzer 5000, serial number **64-001226 (#8)**, was taken out of service for repair on **March 23, 2006**.

I cannot testify that it was working properly and accurately between the dates of **February 23, 2006** and **March 23, 2006**.

This instrument was located at **DUI HQ**.